51" Sink / Storage Unit Installation

MIDMARK®

Special Tools

Level
1/4" deep well socket
Caulking Gun
Belt Sander



DA140000i

DM-OSC051L-0XX (Shown)

Models:

DM-OSC051R-0XX (51" RH Sink / Storage Unit)
DM-OSC051L-0XX (51" LH Sink / Storage Unit)

Use mirror image for right hand units.







DA140100i

NOTES:

• All plumbing must be made by a licensed plumbing contractor and conform to local, city, and national codes pertaining to the installation.

• The wall layout shows a LH unit. If unit being installed is a RH unit, layout and dimensions are a mirror image.

© Midmark Corporation 2003

003-1396-00 Rev. A (8/03)

DRAWER REMOVAL
(If Required for Access)

2" Metal Frame Drawers
or
Writing Surface Assembly



Step 1: Lift or push down release levers (one per side) and slide out drawer assembly

All Molded Drawers



Step 1: Pull drawer outward until it hits stop.

Step 2: Pull sides of drawer outward while lifting to free back of drawer from hooks and remove.

© Midmark Corporation 2003

SINK / STORAGE UNIT INSTALLATION



Step 1: Remove bottom access panels.

Step 2: Adjust leveling screws to fully lowered position and move cabinet to desired position.

Step 3: Level cabinet with leveling screws.

Equipment Alert

Do not adjust leveling screws more then 1 1/2" (3.8 cm) to prevent loss of stability.

© Midmark Corporation 2003

SINK INSTALLATION & PLUMBING CONNECTIONS



Step 1: Install faucet and soap dispenser using instructions included with those items.

Step 2: Install bottom access panel, feeding any lines thru access hole, before connecting Tapmaster™ lines.

Step 3: Refer to the instructions with Tapmaster™ faucet controller for installation procedure.

DA141400i

© Midmark Corporation 2003

DOOR HINGE ADJUSTMENT



Step A: *Side Adjustment*
Turn screw (A) until door is in desired position.

Step B: *Height Adjustment*
Loosen screw (B) approximately 1/4 turn, reposition door, tighten screw.

Step C: *Depth Adjustment*
Loosen screw (C) reposition door, tighten screw.

NOTE:
Doors should be adjusted for the gaps specified on the illustration.

© Midmark Corporation 2003

Midmark Corporation
60 Vista Drive
P.O. Box 286
Versailles, OH 45380-0286
937-526-3662
Fax 937-526-5542
www.midmark.com

MIDMARK®
Because we care.

© Midmark Corporation 2003